

Bid Document/ बिड दस्तावेज़

| Bid Details/बिड विवरण | |
|--|---|
| Bid End Date/Time/बिड बंद होने की तारीख/समय | 17-12-2024 15:00:00 |
| Bid Opening Date/Time/बिड खुलने की तारीख/समय | 17-12-2024 15:30:00 |
| Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से) | 180 (Days) |
| Ministry/State Name/मंत्रालय/राज्य का नाम | Ministry Of Labour And Employment |
| Department Name/विभाग का नाम | Employees State Insurance Corporation |
| Organisation Name/संगठन का नाम | Employees State Insurance Corporation |
| Office Name/कार्यालय का नाम | Esic Hospital Kakinada |
| Total Quantity/कुल मात्रा | 2 |
| Item Category/मद केटेगरी | Anaesthesia Workstation (V2) (Q2) |
| Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) | 49 Lakh (s) |
| OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) | 390 Lakh (s) |
| Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष | 3 Year (s) |
| MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है | Yes |
| Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है | Yes |
| Document required from seller/विक्रेता से मांगे गए दस्तावेज़ | Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |

| Bid Details/बिड विवरण | |
|--|-----------------------------|
| Do you want to show documents uploaded by bidders to all bidders participated in bid?/ | No |
| Past Performance/विगत प्रदर्शन | 50 % |
| Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया | No |
| Comprehensive Maintenance Charges Required | Yes |
| Type of Bid/बिड का प्रकार | Two Packet Bid |
| Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय | 2 Days |
| Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No |
| Evaluation Method/मूल्यांकन पद्धति | Total value wise evaluation |
| Arbitration Clause | No |
| Mediation Clause | No |

EMD Detail/ईएमडी विवरण

| | |
|-----------------------------|---------------------|
| Advisory Bank/एडवाइजरी बैंक | State Bank of India |
| EMD Amount/ईएमडी राशि | 489965 |

ePBG Detail/ईपीबीजी विवरण

| | |
|--|---------------------|
| Advisory Bank/एडवाइजरी बैंक | State Bank of India |
| ePBG Percentage(%) / ईपीबीजी प्रतिशत (%) | 0.01 |
| Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने). | 12 |

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

MEDICAL SUPERINTENDENT

Medical Superintendent, ESIC Hospital, Sambamurthy Nagar, Kakinada -533001
(Medical Superintendent, Esic Hospital, Sambamurthy Nagar, Kakinada-533001)

MII Purchase Preference/एमआईआई खरीद वरीयता

| | |
|---|-----|
| MI Purchase Preference/एमआईआई खरीद वरीयता | Yes |
|---|-----|

MSE Purchase Preference/एमएसई खरीद वरीयता

| | |
|---|-----|
| MSE Purchase Preference/एमएसई खरीद वरीयता | Yes |
|---|-----|

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.
[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

7. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

8. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

9. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Anaesthesia Workstation (V2) (2 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/कमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| Specification | Specification Name/विशिष्टि का नाम | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|---------------------|--|--|
| GENERAL | Product Description | Anesthesia Workstation |
| | Components of Anesthesia workstation | Anesthesia machine , Inbuilt ventilator , vaporisor , Gas delivery system , Multi para monitor |
| Gas Delivery System | Availability of Pin index yokes for gases | Oxygen and Nitrous oxide |
| | NIST /DISS Inlet pipeline with Hoses for O2, N2O and Air | Yes |
| | Type of Flowmeter | Rotameter, Electronic |
| | Mode of gas mixing | Electronic, Pneumatic |
| | Availability of pressure gauges for cylinder and pipelines in anesthesia workstation | Yes |

| Specification | Specification Name/विशिष्ट का नाम | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|-----------------------|--|---|
| | Availability of audible and visual oxygen failure alarms | Yes |
| | Minimum emergency oxygen flow (in liter per minute) | 35 - 70 |
| | Type of hypoxic guard with automatic cutoff of N2O | Electronic |
| | Auxillary Oxygen Flowmeter | Yes |
| | Auxillary Common Gas outlet for use in open circuit | Yes |
| Vaporizer | Number of quick mount type vaporizer for easy interchangeability at a time | 2 |
| | Facility of interlock for vaporizer | Yes |
| | Minimum agent capacity for free volatile anesthetic agent (in ml) | 225 ml |
| | Sensor connection shall be internal to prevent disconnection | Yes |
| | Vaporizer compatible to gases | Sevoflurane, Isoflurane, Desflurane |
| | The vaporizer design should be maintenance free | Yes |
| | Vaporizer calibration requirement | After every 5 years |
| | Should have a drug specific Filling arrangement | Yes |
| Breathing System | Breathing system should have fresh gas de-coupled / compensation , closed circle absorber system | Yes |
| | Number of steps in change over from Mechanical / Spontaneous to bag ventilation | Single step |
| | Number of reusable autoclavable closed silicone circuit supplied | 2 Nos for adult, 2 Nos for pediatric, 2 Nos for Neonatal |
| | Number of Jackson Rees circuit for pediatric patients supplied with workstation | 10 |
| | Breathing system should have leak and compliance test | Yes |
| Anesthesia Ventilator | Available modes of operating ventilator | Manual/spontaneous, Volume controlled, Pressure controlled, SIMV/PCV, SIMV/VCV, CMV |

| Specification | Specification Name/विशिष्ट का नाम | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|---------------|--|--|
| | Ventilator should be capable to pressure support | Yes |
| | Ventilator should be capable to dual mode of ventilation | Yes |
| | Integrated ventilator not require bellows to be change for adults and infants | Yes |
| | Tidal volume of ventilator (in ml) | 10 to 1400 ml |
| | Screen size of Ventilator monitor in inches | 10.1 to 15, 15 .1 and more Or higher |
| | Integrated ventilator shall be able to provide facility to monitor FiO2, dual agent, CO2, Automatic identification of agents, Minimal Alveolar Concentration (MAC) Value | Yes |
| Monitor | Size of monitor in inches | 15.1 to 18, 18.1 and more Or higher |
| | Monitor should not require lengthy start up procedure or calibration for display of parameters | Yes |
| | The graphical and numerical trend with split screen duration | 24 hours |
| | Patient multipara monitor shall be able to display parameters | HR |
| | Capnography to be displayed | Yes |
| | IBP1 and IBP2 to be displayed | Yes |
| | BIS/ENTROPY to be displayed | Yes |
| | ECG monitoring with standard lead configuration, I,II,III, avL, avF&V filters so that it is protected against defibrillator, electro surgery potentials, arrhythmia analysis and ST segment analysis with trends | Yes |
| | NIBP should be determine with oscillometric method, should read systolic, diastolic and mean arterial pressure and pulse rate | Yes |
| | Pulse oximetry with sensors suitable for adults, pediatric & neonate | Yes |

| Specification | Specification Name/विशिष्टि का नाम | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|------------------------|---|--|
| | ETCO2 measurement should be by infra -red absorption technique, measuring mode should be side stream or main stream and response time should be less than 3 seconds | Yes |
| | Monitor should be able to measure and display all parameters in single screen | Yes |
| Alarms | Alarms shall be adjustable to high / low limits with audio and visual alarms | Yes |
| Power supply | Length of main cable in meter with power socket of standard indian make to be provided | ?5 |
| | Power supply | 220-240V AC, 50 Hz single phase |
| Additional Accessories | Transparent re-usable silicon face mask size 0,1, 2, 3, 4 & 5 | 2 each |
| | Breathing bag capacity - 2 Each | 500 ml, 1 Litre, 2 Litre |
| | 3/5 ECG lead | 2 |
| | Nasopharyngeal/rectal and skin Temperature probes should be Provided | 2 Nos Adult, 2 Nos Pediatric |
| | Reusable IBP interface cables | 2 |
| | Disposable IBP transducers | 10 |
| | Accessories for wall mounting | NA (If trolley mounting) |
| ADDITIONAL PARAMETERS | Type of mounting for the Anesthesia workstation | Trolley Mounting |
| | Material used for Anesthesia workstation | Rust-proof ABS plastic with metal reinforcements |
| | Mode of integrated scavenging system with pressure relief valve | Active, Passive |
| | Minimum Battery backup for anesthesia workstation | 1 hour |
| | Facility of manual ventilation in case of electricity and battery failure | Yes |
| | Facility of gas and agent delivery in case of electricity and battery failure | Yes |
| | All the components likes anesthesia machine, integrated ventilator, vaporizer from the same manufacturer | Yes |

| Specification | Specification Name/विशिष्ट का नाम | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|----------------|--|--|
| | The workstation should have one or more drawers to keep accessories, good mobility, anti-static caster wheels with locking facility & conveniently placed handles for easy movement of the machine | Yes |
| | All software updates should be provided free of cost during Warranty period | Yes |
| | The anesthesia machine can be used safely up to 1000 gauss line while being used with MRI | No |
| CERTIFICATIONS | Compliance to Medical Device Rules (MDR) 2017 as amended till date | Yes |
| | Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes |
| | Certification for manufacturing unit | ISO:13485 (Latest) |
| | Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date | Yes |
| | Electrical Protection certification | IEC-60601-1-1 or equivalent BIS |
| | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement | Yes |
| WARRANTY | Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue) | 5 Or higher (year) |

Comprehensive Maintenance

| | |
|------------------------------|--------|
| Warranty of required product | 5 Year |
|------------------------------|--------|

Comprehensive Maintenance

Comprehensive Maintenance Duration (Post Warranty)

5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

| S.No./क्र. सं. | Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी | Address/पता | Quantity/मात्रा | Delivery Days/डिलीवरी के दिन |
|----------------|---|--|-----------------|------------------------------|
| 1 | Arun Kumar Narimeta | 533001,16-24-7/3, Sambamurthy Nagar, Near Port Railway Station, Kakinada | 2 | 15 |

Special terms and conditions-Version:1 effective from 01-07-2024 for category Anaesthesia Workstation (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of

making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.
- 14.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

3. Generic

Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

4. Generic

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 3 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

5. **Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

6. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

7. **Generic**

Manufacturer Authorization:Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

8. **Generic**

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

9. **Generic**

End User Certificate: Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

10. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

11. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

12. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

13. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

14. **Sample Clause**

After award of contract – Successful Bidder shall have to get advance sample approved from buyer before bulk manufacturing / starting bulk supplies. Successful Bidder shall submit

1

samples for Buyer's approval, within 10 days of award of contract. Buyer shall, as per contract specifications framework, either approve the advance sample or will provide complete list of modification

required in the sample within 10 days of receipt of advance sample. Seller shall be required to ensure supply as per approved sample with modifications as communicated by Buyer. If there is delay from buyer side in approval of advance sample – the delivery period shall be refixed without LD for the period of delay in sample approval. In case, the sample is found to have major deviations / not conforming to the Contract specifications, the buyer at its discretion may call for fresh samples for approval before allowing bulk supplies or may terminate the contract after notifying the deviations to the seller.

Unless otherwise provided in the contract, all samples required for test shall be supplied by the contractor free of cost. Where under the contract, the contractor is required to submit an advance sample, any expenses incurred by the contractor on or in connection with the production of stores in bulk, before the sample has been approved unconditionally shall be borne by the Seller and he shall not claim any compensation in the event of such sample being found unacceptable by the Buyer / Consignee.

15. Forms of EMD and PBG

Bidders can also submit the EMD with Banker's Cheque in favour of

MEDICAL SUPERINTENDENT, ESIC HOSPITAL, KAKINADA-533001
payable at
KAKINADA

Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

16. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

MEDICAL SUPERINTENDENT, ESIC HOSPITAL, KAKINADA-533001
payable at
KAKINADA

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

17. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

18. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

19. Certificates

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

20. Certificates

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

21. Certificates

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

22. Turnover

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a

certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

23. **Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

24. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

25. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

26. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export

experience.

11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC. Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C – Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---